

**Monosan HIER Tris-EDTA Buffer pH 9,0(10x), for 1.000 ml**

Reagents

**Instructions for use**

---

|                                |   |
|--------------------------------|---|
| <b>Product name</b>            | Monosan HIER Tris-EDTA Buffer pH 9,0(10x), for 1.000 ml   |
| <b>Intended Use</b>            | HIER T-EDTA Buffer pH 9.0 is a solution developed for heat induced epitope retrieval (HIER) in formalin-fixed paraffinembedded tissue sections on slides. This procedure is primarily used in immunohistochemistry.   |
| <b>Applications</b>            | IHC-P, IHC-Fr, IF   |
| <b>Summary and explanation</b> | Immunohistochemical staining procedures consists of sequential incubation steps with blocking solutions, antibodies and secondary reagents, enzymes and chromogenic substrates carried out on tissue sections. These tissue sections are mostly prepared out of formalin-fixed paraffin-embedded tissue blocks. Cellular structures are very effectively stabilised by formalin fixation which results in optimal morphological preservation of the sample. On the other hand the formalin fixation leads to strong cross-links between proteins. This means that epitopes of antigens are being masked and often are no longer accessible for primary antibodies. In order to enable primary antibodies to bind to antigens the epitopes have to be recovered. Heat induced epitope retrieval (HIER) in buffer solutions of different compositions and pH-values restore structures of the epitopes making them more accessible to specific antibodies. Enzymatic digestion with proteolytic enzymes is another way of recovering epitopes. The primary antibody used determines the appropriate method. |

**FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES**

**Monosan HIER Tris-EDTA Buffer pH 9,0(10x), for 1.000 ml**

Reagents

**Instructions for use**

---

**Principle of method**

HIER T-EDTA Buffer pH 9.0 is a 10fold concentrated EDTA solution in Tris buffer with additives of detergent and stabilising substances. For preparation of the working strength solution the buffer concentrate is diluted 1:10 with deionised or distilled water. The resulting solution has a pH of 9.0 (8.8 to 9.2). HIER T-EDTA Buffer pH 9.0 is a very efficient epitope retrieval solution in immunohistochemical staining procedures to be used with primary antibodies of many different specificities. It leads to considerably stronger signals compared with usually used citrate buffer.

**Reagents provided**

100 ml HIER T-EDTA Buffer pH 9.0 (10fold concentrated, adequate for 1 litre ready-to-use T-EDTA Buffer)

**FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES**

---

**Instructions for use**

---

**Storage and handling**

The solution should be stored at 2-8°C. Do not freeze it. Under these conditions the solution is stable up to the expiry date indicated on the label. Do not use product after the expiry date. If stored at room temperature the solution is stable for at least 10 months from the date of delivery. The prepared working strength solution is stable for 1 month, if stored at 2-8°C. A positive and a negative control have to be carried out in parallel to the test material. If you observe unusual staining or other deviations from the expected results which could possibly be caused by this reagent, please contact our technical support.

**Reagent preparation**

Preparation of the T-EDTA buffer working strength solution:

- Dilute HIER T-EDTA Buffer concentrate 1:10 with deionised or distilled water and mix thoroughly.
- The pH-value should be at 9.0 (8.8 to 9.2). If necessary adjust pH-value with diluted NaOH or HCl solution.

**Procedure**

HIER T-EDTA Buffer is suitable for various HIER-methods such as steamer, pressure cooker, autoclave, water bath, and microwave oven. Tissue sections used in heat induced epitope retrieval should always be placed on adhesive slides. Epitope retrieval is carried out after dewaxing and rehydration of the sections. Exemplary protocol using steamer: 1. Prepare the working strength solution by diluting the buffer concentrate as described above and transfer to a Coplin jar. Please make sure that there is enough volume to cover the tissue sections on the slides completely. 2. Fill steamer with water according to instruction manual, close lid and start. 3. After 10 minutes place Coplin jar with T-EDTA buffer in the steamer, close the lid and heat the solution for 20 minutes. 4. Place slides with tissue sections into the preheated solution and close the lid. Tissue sections have to be completely covered with T-EDTA buffer solution. 5. Incubate slides 20 – 40 minutes. The optimal incubation time needs to be elaborated by the operator. 6. After the incubation take the Coplin jar with slides out of steamer and let cool down at room temperature for about 20 minutes. 7. Remove T-EDTA buffer, rinse slides with wash buffer and proceed with immunohistological staining.

**Expected results**

During the reaction of the substrate with horse radish peroxidase or alkaline phosphatase in the presence of a chromogen, a coloured precipitate is formed at the location of the bound primary antibody. This reaction only takes place if the target antigen is existent in the tissue. The chromogen used determines the colour of the precipitate. The analysis is carried out using a light microscope.

**Instructions for use**

---

**Trouble shooting**

If you observe unusual staining or other deviations from the expected results please read these instructions carefully, or contact our technical support. Also refer to the instructions of the detection systems for guidance on general troubleshooting.

**FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES**

**Instructions for use**

---

**Quality control**

We recommend carrying out a positive and a negative control with every staining run. The positive control permits the validation of appropriate processing of the sample. If the negative control has a positive result, this points to unspecific binding. Please refer to the instructions of the detection system for guidance on general quality control procedures.

**Performance**

Studies have been conducted to evaluate the performance of the kit reagents. The product has been found to be suitable for the intended use

**Limitations of procedure**

Immunohistochemistry is a complex technique involving both histological and immunological detection methods. It requires a highly trained histotechnologist. Tissue processing and handling prior to immunostaining, for example variations in fixation and embedding or the inherent nature of the tissue can cause inconsistent results (Nadji and Morales, 1983). Inadequate counterstaining and mounting can influence the interpretation of the results. Sanbio guarantees that the product will meet all requirements described from its shipping date until its expiry date, as long as the product is correctly stored and utilized. No additional guarantees can be given. Under no circumstances shall Sanbio be liable for any damages arising out of the use of the reagent provided.

**Precautions**

Use by qualified personnel only. Wear protective clothing to avoid contact of reagent and specimen with eye, skin or mucous membranes. In case of reagent or specimen coming into contact with a sensitive area, wash area with large amounts of water. Microbial contamination of the reagent must be avoided, since otherwise non-specific staining may occur. ProClin 300 is used for stabilisation. A material safety data sheet (MSDS) is available upon request.

**References**

1. Miller RT et al. Appl Immunohistochem Mol Morphol 8:228-235, 2000
2. Nadji M and Morales AR Ann N.Y. Acad Sci 420:134-139, 1983
3. -

**FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES**